

NANO ZS



Size Distribution Report by Volume

Sample Details

Sample Name: JM 9.2.05

SOP Name: CSL-00.sop

General Notes:


File Name:	CSL-size.dts	Dispersant Name:	Water
Record Number:	447	Dispersant RI:	1.33
Material RI:	1.25	Viscosity (cP):	0.8872
Material Absorbtion:	0.01	Measurement Date and Time:	Wednesday, September 07, ...

System

Temperature (°C):	25.0	Duration Used (s):	60
Count Rate (kcps):	231	Measurement Position (mm):	4.65
Cell Description:	Disposable sizing cuvette	Attenuator:	10

Results

	Diam. (nm)	% Volume	Width (nm)
Polydispersity Index: 0.614	Peak 1: 2.07	96.5	0.2
	Peak 2: 20.83	3.5	4.5
	Peak 3: 177.14	0.0	53.1



Colloidal Science Laboratory, Inc.

Westampton, NJ 609-267-2065